















## Clinical Confidence




## Precision

When your focus is on reducing dose and obtaining excellent image quality, FDR D-EVO is for you.




### Patented ISS capture technology promotes high sensitivity

The FDR Clinica Series of Components is intended for use with FDR D-EVO detectors featuring Fujifilm's proprietary Irradiated Side Sampling (ISS) technology, which positions its capture electronics (TFTs) at the irradiation side, in contrast to traditional detectors. This design significantly suppresses scattering and attenuation of X-ray signals, improving efficiency to produce sharper images at lower doses compared to traditional designs.\*







ISS system reading technology

# Optimal Efficiency



The FDR Clinica Series of Components has all the advanced features you need to enhance workflow and improve the patient experience.


## Utility


New features improve workflow and patient experience.



### 1 Tilting Wall Stand

A wide variety of exams can be performed using the optional tilting wall stand. Nonambulatory patients can be easily imaged in the seated position.





#### 2 Dual-reference Rotation

Both top- and center-reference rotation are available in our rotating tray, with automatic connection to the detector. The detector can be rotated while mounted in the tray with no need to remove and remount, greatly reducing handling.





#### 3 Automatic Connection

The detector cable and connector are built into the tray, enabling the detectors to be automatically connected when set in the tray. This new function greatly improves efficiency and provides a clean look with no need to manage cabling. The automatic connection to the detector keeps the battery charged in the event that wireless operation is required.

## Advanced image processing: Dynamic Visualization





Conventional image



New image




Customized image

Fujifilm's renowned digital image processing expertise, including Dynamic Visualization, optimizes image display based on recognition of contrast levels throughout the entire exposure field. The resulting first-up images have outstanding detail and greater window and leveling capability in PACS.

## Versatility


The FDR Clinica Series of Components provide flexibility, with a wide selection of power ratings to meet your requirements.



Luminosity

Light Source

Standard



### FDR Clinica Series of Components Specifications

	FDR Clinic	ca GXR Series of Co	omponents		
		X-RAY GENERATOR			
	Capacitor Type	Line-power Type			
Product Name	GXR-C40	GXR-40	GXR-52	GXR-68	
Output Rating	40 kW	40 kW	52 kW	68 kW	
Line Nominal, Phase	110/120VAC, 230VAC, 1Φ	230 VAC, 1Φ 400 VAC, 480 VAC, 3Φ	400VAC, 480VAC, 3Φ		
Line Voltage Range	±10% (Frequency: 50/60 Hz)				
kV Range			40-150 kV	40–150 kV, 1 kV step	
mA Range	10 to 500 mA	10 to 500 mA	10 to 640 mA	10 to 800 mA	
Timer Range	0.001 to 10 sec, 38 steps				
mAs Range	0.1 to 500 mAs				
Max. Power Output	500mA@80kV 400mA@100kV 320mA@125kV	500 mA@80 kV 400 mA@100 kV 320 mA@125 kV	640 mA@81 kV 500 m A@104 kV 400 mA@130 kV 320 mA@150 kV	800mA@85kV 640mA@106kV 500mA@136kV 400mA@150kV	
Minimum Breaker Rating	15A (230VAC, 1Φ) 30A (110VAC, 3Φ)	100A (230VAC, 1Φ) 65A (400VAC, 3Φ) 50A (480VAC, 3Φ)	75 A (400 VAC, 3Ф) 65 A (480 VAC, 3Ф)	90A (400VAC, 3Ф 75A (480VAC, 3Ф	
Anatomical Programs	User-programmable max. 1,280 programs with APR utility software				
Technique Selection	4 point display (kV, mA, Time, mAs)				
		X-RAY TUBE			
Focal Spot Size	0.6/1.5 mm		0.6/1.2 mm		
Max. Anode HU	200 kHU	* * * *		300 kHU	
Target Angle	14° 12°				
		COLLIMATOR			
Field Shape	Rectangular				
Max. Field Size	More than 17x 17 in at 100cm SID				
Inherent Filtration	Min. 2.0 mmAl eg.				

Over 160 LUX at 100 cm SID (Typ. 250 LUX)

Single LED

Laser line, Tape measure, Rotating flange

#### lacktriangle Continued from the previous page.

Patient Table				
			4-way Floating Tabletop	6-way Elevating Table
Movement	Tabletop	Longitudinal	39 (±20) in	
	Tabletop I	Transversal	10 (±5) in	
		Travel		11 (22-33) in
	Vertical	Speed		17 mm/sec
	vertical	Operating		Motorized movement by footswitch DC-motor (Linear Actuator)
Bucky		Longitudinal	14 in with standard tray	
		Inherent Filtration	1.2 mmAl at 100kV	
		Max. Patient Weight	660 lbs (300 kg )	
		Size	87 (W) x 30 (D) x 2 (H) in	87 (W) x 32 (D) x 2 (H) in
Bucky Type/Grid		Fixed	FD 100 cm, 103 lpi, ratio 10:1	
Lock (Brake)			EM Lock, beam sensor on/off	EM Lock, Foot Switch on/off
Center Indication			Buzzer sound and LED	Transverse center, height center
Dimension/Weight			87 (W) x 30 (D) x 26 (H) in 330 lbs (150 kg)	87 (W) x 32 (D) x 33 (H) in 573 lbs (260 kg)

Vertical Wall Stand				
Cassette Stroke	Vertical 65 in (17-81 in from floor to Bucky center)			
Bucky Type/Grid	Fixed	FD 150 cm, 103 lpi, ratio 10:1		
Lock (Brake)	EM Lock, Switch on/off			
Balance	Counter Weight			
Dimension/Weight	Max. 85 (H) x 26 (W) x 16 (D) in / 264 lbs (120 kg)			

Floor-mounted Tube Stand				
Tube Rotation Angle	±135°			
Tube Stroke	Longitudinal	Max. 87 in		
	Lateral	7 in		
	Vertical	62 in (17-80 in from floor to focus)		
Lock (Brake)	EM Lock, Foot Switch on/off			
Balance	Counter Weight			
Column Rotation	90° step			
Dimension/Weight	91 (H) x 54 (D) x 118 (W) in / 529 lbs (240 kg)			

### DR-ID 600 ( € 0123

#### EDR D EVO for Digital Padiogram

(DR-ID 600, including DR-ID 300CL)

FDR D-EVO G35i and C35i are available for the FDR Clinica Series of Components.
For details, please refer to the FDR D-EVO Series brochures.




FDR D-EVO C35i



FDR D-EVO G35i

#### Dimensions



Specifications are subject to change without notice.

